| NITRATE ELECTRODE METHOD  SM 4500-NO <sub>3</sub> D-2000 (2011)  ADDITIONAL QC REQUIREMENTS FOR THIS METHOD: Certified or Accredited laboratories using this method are assessed to applicable requirements of SM 1020 and SM 4020. |                                      |                   |          |     |          |  |
|---|--------------------------------------|-------------------|----------|-----|----------|--|
| Facility Name:  | VELAP ID                             |                   |          |     |          |  |
| Assessor Name:Analyst Name:   | Inspection Date                      |                   |          |     |          |  |
| Relevant Aspect of Standards  | Method<br>Reference                  | Y                 | N        | N/A | Comments |  |
| Records Examined: SOP Number/ Revision/ Date  |                                      |                   | Analyst: |     |          |  |
| Sample ID: Date of Sample Preparation:  |                                      | Date of Analysis: |          |     |          |  |
| Were unpreserved samples stored as follows?  Nonpotable: ≤ 6°C up to 48 hours  Drinking water: 4°C up to 48 hours unless chlorinated, which can be held up to 14 days   | 40CFR 136.3,<br>40CFR 141.23         |                   |          |     |          |  |
| When NO <sub>3</sub> and NO <sub>2</sub> where determined as separate species, were samples never acidified?  | NO <sub>3</sub> <sup>-</sup> A.2     |                   |          |     |          |  |
| Did buffer solution contain Ag <sub>2</sub> SO <sub>4</sub> , sulfamic acid, Al <sub>2</sub> (SO <sub>4</sub> ) <sub>3</sub> , and a buffer at pH 3?  | NO <sub>3</sub> <sup>-</sup> D.3.d   |                   |          |     |          |  |
| Was the pH meter capable of 0.1 mV resolution?  | NO <sub>3</sub> <sup>-</sup> D.2.a   |                   |          |     |          |  |
| Was the outer chamber of the reference electrode filled with (NH <sub>4</sub> ) <sub>2</sub> SO <sub>4</sub> ?  | NO <sub>3</sub> <sup>-</sup> D.2.b   |                   |          |     |          |  |
| Were the manufacturer's instructions followed for the care and storage of the nitrate ion electrode?  | NO <sub>3</sub> <sup>-</sup> D.2.c   |                   |          |     |          |  |
| Was 10 mL buffer added to 10 mL of sample/standard in a 50 mL beaker and stirred with a magnetic stirrer?   | NO <sub>3</sub> <sup>-</sup> D.4.a/b |                   |          |     |          |  |
| Were electrode tips allowed to stabilize in solutions about 1 minute prior to the recording of millivolt readings?  | NO <sub>3</sub> <sup>-</sup> D.4.a/b |                   |          |     |          |  |
| Were electrode tips rinsed and blotted dry between samples?   | NO <sub>3</sub> <sup>-</sup> D.4.a   |                   |          |     |          |  |
| Were calibration curves made by semilogarithmic plots with concentration on the logarithmic axis and millivolt responses on the linear axis?  | NO <sub>3</sub> <sup>-</sup> D.4.a   |                   |          |     |          |  |
| Were calibration slopes +57 ± 3 mV/decade at 25°C?  | NO <sub>3</sub> <sup>-</sup> D.4.a   |                   |          |     |          |  |
| Were samples and standards measured at approximately the same temperature?  | NO <sub>3</sub> <sup>-</sup> D.4.b   |                   |          |     |          |  |
| Notes/Comments:   |                                      |                   |          |     |          |  |